


THESE DRAWINGS ARE THE PROPERTY OF QUALTEK ELECTRONICS AND NOT TO BE REPRODUCED FOR OR COMMUNICATED TO A THIRD PARTY WITHOUT THE EXPRESS WRITTEN PERMISSION OF AN OFFICER OF QUALTEK ELECTRONICS.

| REVISION |             |      |          |
|----------|-------------|------|----------|
| REV.     | DESCRIPTION | DATE | APPROVED |
|          |             |      |          |
|          |             |      |          |
|          |             |      |          |



NOTES:

1. CORDAGE: SJT, 16AWG / 3C, UNSHIELDED, NA COLOR CODE, TEMP. RATING: 60°C, RATING: 125V, 13A, JACKET COLOR: GRAY
2. PLUG: NEMA 5-15P HOSPITAL GRADE (CLEAR)
3. CONNECTOR: IEC 60320 C13 (GRAY)
4. APPROVALS: UL, CSA
5. ROHS COMPLIANT
6. CAUTION LABEL

**CONTROLLED**

|                |                    |   |                |  |
|----------------|--------------------|---|----------------|--|
| Dim range (mm) | Tolerance +/- (mm) | Qualtek Electronics Corp.<br>PPC DIVISION |                |  |
|                |                    | Part Number:                              | 233086-06      | <span style="border: 1px solid green; padding: 2px;">RoHS Compliant</span> |
| 0-1            | 0.15               | Description: POWER CORD HOSPITAL GRADE    |                |  |
| 1-6            | 0.20               |   |                |  |
| 6-18           | 0.30               |   |                |  |
| 18-50          | 0.50               | UNIT: in                                  |                |  |
| 50-120         | 0.70               | Drawn / Date                              | Checked / Date | Approved / Date  |
| 120-250        | 0.90               | R.Nuti<br>11-12-08                        | 11-12-08       | LY / 11-12-08  |
| 250-500        | 1.20               |   |                |  |